

SA-E-6200

Saliva Estradiol ELISA kit

This Saliva Estradiol ELISA kit is designed to measure free estradiol levels in human saliva within 3 hours, supporting both research and diagnostic efforts in areas such as menstrual cycle monitoring, pregnancy assessment, menopause evaluation, endocrine disorder investigation, and the study of hormone-dependent cancers.



Sample type	Saliva
Capacity	96-well plate
Sensitivity	0.716 pg/mL
Range	0/1 - 100 pg/mL
Assay time	3h
Reactivity	Human
Application	Research use or in vitro diagnostics (IVD in EU-only)
Storage	Store at 2-8°C for to 6 months
Sample collection & storage	Saliva: collect 5 samples within a period of 2 hours (multiple sampling), before meal. Store samples at 2-8°C for up to one week. For longer periods, store at -20°C.
Sample preparation	Sample incubation (30 min), Enzymatic conjugation (120 min)
ELISA	Incubation, revelation and read steps (30 min)

To see data, publications, and download instructions for use and MSDS, visit:
<https://www.immusmol.com/product/saliva-estradiol-elisa-kit/>